

MICHIGAN DEPT OF AGRICULTURE & RURAL DEVELOPMENT  
 LABORATORY DIVISION  
 WEIGHTS AND MEASURES PROGRAM  
 (517) 655 - 8202  
 michigan.gov/wminfo


Device Detail


**Insp Date:** 11/16/2017 **Business ID:** 91169  
**Business:** ENVIRO SOLIDS  
 6011 WYOMING  
 DEARBORN, MI 48126

**Inspection:** BB001339  
**Store ID:**  
**Phone:** (313)582-8032  
**Inspector:** 015 James Barnes  
**Reason:** FIELD AUDIT

**Make:** FAIRBANKS **Type:** Heavy Capacity Scale **Results:** Sealed  
**Model:** FB 2550 **Subtype:** 09-5 (25000) **Other:**  
**Serial:** 140870050235 **Insp Type:** Maintenance **Seal #:** AUDIT TRAIL  
**Loc:** RIGHT OF ENTRANCE  
**Capacity:** 200,000 **Units:** lb  
**Class:** IIII **Sections:** 5  
**NTEP:** Yes **Weighing Elements:** Electronic Load  
**Elapsed Time:** 2.25  
**d:** 20  
**CLC or Section Capacity:** 90,000  
 Cel  
**Indicating Elements:** DWI **Recording Elements:** CPU Interface **Deck Size:** 11' X 80'  
**Invoice#:** N/A

Type	Name	Inc/Base	Actual	Display	Err	Tol	Res	Notes
Test	Zero Load	0	0	0	0	20	Pass	
Test	Increasing Load	5000	5000	5000	0	20	Pass	Section 1
Test	Increasing Load	4000	9000	9000	0	20	Pass	
Test	Increasing Load	4000	13000	13000	0	40	Pass	
Test	Shift	0	13000	13000	0	40	Pass	Section 2
Test	Shift	0	13000	13020	20	40	Pass	Section 3
Test	Shift	0	13000	13000	0	40	Pass	Section 4
Test	Shift	0	13000	13000	0	40	Pass	Section 5
Test	Range(Shift)	0	13000		20	40	Pass	
Test	Increasing Load	4000	17000	17000	0	40	Pass	Section 1
Test	Increasing Load	4000	21000	21000	0	60	Pass	
Test	Increasing Load	4000	25000	25000	0	60	Pass	
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 2
Test	Shift	0	25000	25020	20	60	Pass	Midspan
Test	Shift	0	25000	25040	40	60	Pass	Section 3
Test	Shift	0	25000	25040	40	60	Pass	Midspan
Test	Shift	0	25000	25020	20	60	Pass	Section 4
Test	Shift	0	25000	25020	20	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 5
Test	Range(Shift)	0	25000		40	60	Pass	
Test	Decreasing Load	-12000	13000	13000	0	40	Pass	Section 1
Test	Zero Load	0	0	0	0	20	Pass	
Test	Strain	28020	25000	53020	0	60	Pass	

  
 Manager



ps, Lab & Operations

## Device Detail

**Notes:**

Reviewed service company documents onsite from Grand Rapids Metrology dated 8-14-17.

**AUDIT TRAIL**

Calibration Date: 8-16-16

Calibration Count: 21

Configuration Date: 1-12-15

Configuration Count: 2

Any questions or concerns regarding this inspection please contact Inspector Jim Barnes at (517)449-6241.



Manager



Environmental Services, Lab & Operations